



Product Service

CERTIFICATE

No. Z1A 05 06 10992 005

Holder of Certificate: Grimm Aerosol Technik GmbH & Co. KGDorfstr. 9
83404 Ainring
GERMANY**Factory(ies):** 24138**Certification Mark:****Product:** Meßgeräte für Stäube, Rauch, Nebel**Model(s):** 107, 1.108, 1.109

Parameters:	Rated voltage:	18 V DC Power supply or 12 V DC Accumulator
	Rated current:	0,75 A
	Protection class:	III
	Laserclass:	1

Tested according to: EN 61010-1:2001
EN 60825/A2:2001

The product meets the requirements of the German Equipment and Product Safety Act. The Certification marks shown above can be affixed on the product. The use of the GS-Mark is permitted until the listed date, the use of the TÜV-Mark is unlimited, unless it is cancelled. See also notes overleaf.

Test report no.: 028-70093574-000**GS-Mark valid until:** 2010-06-15**Date,** 2005-06-15

Page 1 of 1

